

Wiring

Internal electrical TPA2 connection resembles simple switch with normally connected contacts. After stylus is actuated, contacts become disconnected. Signals are isolated from instrument body.

	Function
	Not used (unconnected)
	Signal
	Signal
	Not used (unconnected)

Revision #5
Created 15 November 2020 08:54:26 by Saulius
Updated 4 May 2021 14:23:35 by Saulius